

6 Installation Planning: Checklist

In order to avoid any unnecessary delays during the installation and the machine start-up, the following points of the checklist below should be carried out prior to the installation.
Check and discuss all the required measures for the installation by means of this checklist. Remarks on the individual items may be made on the back of the list.

We ask you to give this checklist to your local A-G representative then.

Customer: .....

Department: .....

System components: .....

.....

.....

.....

.....

Desired installation date: .....

Signed: .....
Signature

Remarks:

p.t.o.

6.1 Checklist

	Required external connections		Prepared: yes	no
1.	**ID Station**	3 outlets for PC, ID tablet and monitor	☐	☐
		Ethernet connection (twisted pair) prepared, if RIS LINK will be installed	☐	☐
2.	**Client Station**	Ethernet connection (twisted pair) prepared	☐	☐
		2 outlets for PC and monitor	☐	☐
3.	**ADC QS Server Station**	Outlets • minimum 2 outlets for DELL Precision 420	☐	☐
		Ethernet connection (MAU or Twisted pair) prepared	☐	☐
4.	**D**irect **R**emote **A**ccess	Remote access via Service Host must be guaranteed! (telephone line required)	☐	☐
5.	**Hardcopy printer available via network**	See separate Installation Planning Instructions	☐	☐
	General			
		Transport to the installation site, unpacking and removal of the devices from the pallet must be done by the carrier.		
6.	For the connection of the system components to the local network please plan an appointment concerning network configuration with the local network administrator.		☐	☐

7 Form to fill in the Network Parameters

Digitizer	Example	1st Digitizer	2nd Digitizer	Remarks
hostname	adcc1			
ip_addr.	192.9.200.199			
Subnet_mask				
default router				
AE_title	ADCC1			
Station Name *	ADCC1			

ADC QS Server Station	Example	1st Server Station	2nd Server Station	Remarks
hostname	adc_qs			
ip_addr.	192.9.200.202			
subnet_mask				
default router				
AE_title	ADC_QS			
Station Name *	ADC_QS			

ID - Station	Example	1st ID-Station	2nd ID-Station	Remarks
hostname	id207			
ip_addr.	192.9.200.207			
Subnet_mask				
default router				
AE_title	ID207			
Station Name *	ID207			

= main components

ADC QS Client Station	Example	1st Client Station	2nd Client Station	Remarks
hostname	qc206			
ip_addr.	192.9.200.206			
Subnet_mask				
default router				

Hard Copy Printer	Example	1st HCP	2nd HCP	Remarks
hostname	mg1			
ip_addr.	192.9.200.201			
Subnet_mask				
default router				
AE_title	ADC_LR1			
Station Name *	ADC_LR1			

PACS Archive Station	Example	1st Archive Station	2nd Archive Station.	Remarks
hostname	simas1			
ip_addr.	192.9.200.101			
Subnet_mask				
default router				
AE_title	SIMAS1			
Station Name *	IMPAX			

*) Friendly name (Station name) appears in the USER interface

Site filled in by Date

AGFA and the Agfa-Rhombus are trademarks of Agfa-Gevaert AG, Germany

Herausgegeben von / Published by / Edité par / Pubblicato da / ..... / Editado por

Agfa-Gevaert AG
Fototechnik
Tegernseer Landstraße 161
D - 81539 München

Printed in Germany / Imprimé en Allemagne /
stampato in Germania / .. / Editado por

AGFA